


校准证书

CALIBRATION CERTIFICATE

证书编号:



第 1 页 共 3 页

Certificate No.

J202101282113-0007

Page of

委托方

Client

联络信息

Contact Inf.

See more at: <http://www.china-gauges.com/>

仪器名称

Description

Universal Spring Hammer (IK01~IK06)

型号/规格

Model/Type

LX-T06

制造厂

Manufacturer

HK LEE HING INDUSTRY CO., LIMITED

出厂编号

Serial No.

JLX2021G200671

管理号

Asset No.

接收日期

Receipt Date

2021年01月28日

Y M D

校准日期

Calibration Date

2021年02月02日

Y M D

发布日期

Issued Date

2021年02月02日

Y M D

批准

Approved by

李平 (副总工)

审核

Inspected by

李俊峰

校准

Calibrated by

林屹涛



总部地址(Headquarters Add):广东省广州市黄埔大道西平云路163号

No.163.Pingyun Rd, West of HuangPu Ave.Guangzhou.Guangdong .China

承校实验室地址(Add. of the Cal. Lab): 广东省广州市黄埔大道西平云路163号

No.163.Pingyun Rd,West of HuangPu Ave.Guangzhou,Guangdong,China

联系电话(Tel.):020-38699960,66830999,400-602-0999

传真(Fax):020-38695185

邮政编码(Postcode):510656

网站(Website):[http:// www.grgtest.com](http://www.grgtest.com)电子邮件(E-mail):grgtest@grgtest.com

扫一扫验真伪

校准说明

DIRECTIONS OF CALIBRATION

证书编号: J202101282113-0007

第 2 页 共 3 页

Certificate No.

Page of

- 1.本实验室的质量管理体系符合ISO/IEC 17025:2017标准的要求。
(The quality system is in accordance with ISO/IEC 17025:2017.)
 - 2.本结果仅对本次校准样品有效。未经实验室批准,不得部分复制。如有疑问请在15个工作日内反馈。
(The result is only valid for the calibrated sample.The certificate shall not be reproduced except in full,without the written approval of our laboratroy .please feedback to us within 15 days if you have any question.)
 - 3.本证书编号具有唯一性,后缀若带有“-Gx”的证书为替换证书,自发出后原证书即刻作废。
(Each certificate has a unique number. The suffix of "-Gx" will be added to the number as a replacement of the old version. The original certificate will be officially invalid once the new certificate number is issued.)
 - 4.证书中如有最大允许误差、判定结果,仅供参考,其中“P”代表“合格”,“F”代表“不合格”。使用人员还应结合实际测量要求,评估校准结果测量不确定度对符合性评定的影响。(MPE & judgement result in the datasheet is only for reference, "P" represents "Pass" and "F" represents "Fail".Whereas users should evaluate the effects of MU of calibration results on conformity determination associated with actual measurement.)
 - 5.本次校准的技术依据及CNAS认可范围,超出范围的内容未被认可。详细认可范围请查看CNAS网站中注册编号为L0446的证书附件。(Reference document and accredited scope by CNAS for calibration, beyond which isn't accredited. Please see the attachment of certificate No.L0446 on CNAS website for details.)
- JJF 1475-2014 弹簧冲击器校准规范(C.S. for Spring Hammers) 长度: (1~50)mm 力值: (1~50)N 能量: (0.2~2) J

6. 本次校准使用的主要测量标准(Main Standards of Measurement Used in the Calibration.):

名称 / 型号	编号	证书号/有效期	溯源机构	技术特征
Description / Model	Serial No.	Certificate No./ Due Date	Traceability Institute	Technique Character
弹簧冲击器校准装置/CJ-XZ	CJ010207	J202006046177- 0002 2021-06-06	广州广电计量检 测股份有限公司	$U = (0.005 \sim 0.01) J (k=2)$
弹簧冲击器专用码/10N	115300	J202003191938- 0003 2021-03-19	广州广电计量检 测股份有限公司	MPE:±0.2N
数显卡尺/(0~200)mm	K15D078852	J202008038612- 0007 2021-08-03	广州广电计量检 测股份有限公司	MPE:±0.03mm
影像测量仪/VMS-2010F	VJC50083	J202003191938- 0005 2021-03-22	广州广电计量检 测股份有限公司	MPE: ± (2.5+L/100) μm

7. 校准地点、环境条件(Place and environmental conditions of the calibration):

地点	广州计量力学室	温度	23	℃	相对湿度	46	%
Place		Temperature			Relative Humidity		

8. 建议复校时间间隔: 1年,送校单位也可按实际使用情况自主决定。
Suggested calibration interval is 1 year or it can be altered depending on the actual usage of the user.

校 准 结 果
RESULTS OF CALIBRATION

证书编号: J202101282113-0007

第 3 页 共 3 页

Certificate No.

Page of

1、外观及各部分相互作用: 正常

Appearance and the function of each part of interaction:

2、锤头球面半径校准: 符合要求

技术要求: (10±0.2)mm 结论: P

Calibration of Radius of sphere

Technical requirements

Conclusion

3、释放撞击元件所需的力: 符合要求

技术要求: ≤10N 结论: P

Calibration of The force required to release the impact element

Technical requirements

Conclusion

4、锤弹簧未压缩状态时, 锤头顶点缩进释放锤头端面的距离:

Calibration of Distance 符合要求

技术要求: >1mm 结论: P

5、冲击能量示值误差校准:

Technical requirements

Conclusion

Calibration of Impact energy:

标称值(J)	实测平均值(J)	示值误差(J)	允许误差(J)	结论 (P/F)
Nominal Value	Average Value	Error	MPE	Conclusion
0.14	0.131	+0.009	± 0.014	P
0.2	0.190	+0.010	± 0.02	P
0.35	0.341	+0.009	± 0.03	P
0.5	0.499	+0.001	± 0.04	P
0.7	0.678	+0.022	± 0.05	P
1	0.977	+0.023	± 0.05	P

注: 0.14J不在CNAS认可范围内

备注:

Notes:

结论(Conclusion): 所校项目符合技术要求

1.本报告中的扩展不确定度是由标准不确定度乘以包含概率约为95%时的包含因子 k 。
The expanded uncertainty is given in the report by the standard uncertainty multiplied by the probability of about 95% when the factor k .

1.1 冲击能量: $U_{rel}=3.0\%$ ($k=2$)

2.依据(Reference document)

JJF 1059.1-2012 测量不确定度评定与表示

(JJF 1059.1-2012 Evaluation and Expression of Uncertainty in Measurement)

(以下空白)

(The below is blank)